



Precision Polymer Engineering



Reference: USP88/121-PPE-Perlast
Date: 24th June 2026

Certificate / Bescheinigung / Certificat

USP <88> Biological Reactivity Tests *in vivo*

E70Q, S60H, S70C, S71U, S80U, XS7H,
G74S, V70WF, V70SW, V75RF, V75SR

This is to certify that the compounds listed above conform to the United States Pharmacopeia USP <88> for the Biological Test for Plastics, Class VI at 121°C.

Wir bestätigen hiermit, daß die oben aufgeführten Werkstoffe den United States Pharmacopeia USP <88> bezüglich Biological Test for Plastics, Klasse VI bei 121°C entsprechen.

Nous certifions par la présente que nos mélanges mentionnés ci-dessus sont conformes à la norme United States Pharmacopeia USP <88> Biological Test for Plastics, Classe VI à 121°C.

Stefano Arrigoni
Head of Research and Development
IDEX Sealing Solutions

Precision Polymer Engineering Limited
Greenbank Road, Blackburn, BB1 3EA, England | Tel: +44 (0) 1254 295400

Precision Polymer Engineering LLC
3201 S. Bluebell Road, Brenham, TX 77833, USA | Tel: +1 979 353 7350

www.prepol.com

Precision Polymer Engineering is a unit of IDEX Corporation